



中国检验检疫科学研究院  
综合检测中心



中国认可  
国际互认  
检测  
TESTING  
CNAS L0424

# 检验检测报告

## INSPECTION AND TEST REPORT

报告编号: CAIQS0026002463003

样 品 名 称	0A李斯特氏菌显色培养基平板
检 测 类 别	委托检测
委 托 单 位	无锡市赛微生物技术有限公司
签 发 日 期	2026年05月22日

中国检验检疫科学研究院综合检测中心

CHINESE ACADEMY OF INSPECTION AND QUARANTINE COMPREHENSIVE TEST CENTER

# 重要声明

## Important Statement

1. 本检验检测报告由报告封面和报告内容组成。无报告封面，以及报告内容不完整的，报告无效。

This test report consists of cover and contents. If the report has no cover or the contents of the report are not complete, the report shall be invalid.

2. 检验检测报告封面或报告结论或骑缝位置处无本单位检验检测专用章、复印件未重新加盖检验检测专用章、签字不完整、涂改或增删的，报告无效。

If the cover, conclusion or cross-page place of the report is not stamped with the special seal for test of CAIQTEST, or if any copy is not re-stamped with the special seal for test, or if the signature is incomplete, altered, increased or deleted, this report shall be invalid.

3. 本报告中委托方信息、样品信息均由委托方提供并确认，由委托方承担证实信息准确性、完整性的责任。

The information of client and samples in this report is provided and confirmed by the client; the client is responsible to verify the accuracy and completeness of such information.

4. 委托方自行送检的样品，检测结果仅适用于收到的样品，与被测样品取样的来源无关。

The test results are only applicable to the samples as submitted by the client, and are unrelated to the source of tested samples.

5. 未加盖资质认定标志（CMA）的检验检测报告，不具有对社会的证明作用，仅作为科研、教学或内部质量控制之用。

The test report without qualification certification logo (CMA) shall have no function of proof to the society, and may be only used for scientific research, teaching or internal quality control.

6. 未经本单位批准，不得复制（全文复制除外）报告。

Without the approval of CAIQTEST, this report may not be reproduced (except as reproduced in whole).

7. 未经本单位书面同意，不得将此报告用于商业性及不当宣传。

Without the written consent of CAIQTEST, this report may not be used for commercial or improper propaganda.

8. 对检验检测报告有异议的，应当自收到检验检测报告之日起十五日内向本单位书面提出，未提出、逾期提出、未以书面形式提出的，均视为无异议。

If you have any objection about the test report, please put forward the objection to CAIQTEST within fifteen days after the receipt of the report. If you do not put forward, fail to put forward as scheduled, or fail to put forward in writing, you will be deemed to have no objection.

地址：北京市朝阳区高碑店北路甲3号

Address: No. A3, Gaobeidian North Road, Chaoyang District, Beijing

邮编：100123

Postcode: 100123

电话：400-987-0567

Tel: 400-987-0567



检验检测报告  
INSPECTION AND TEST REPORT

报告编号 (No.) : CAIQS0026002463003共1页, 第1页    page 1 of 1

委托单位名称 Client		无锡市赛微生物技术有限公司						
委托单位联络信息 Client Contact Information		无锡市北塘区光电新材料科技园会北路28-95						
样品名称 Sample Name		0A李斯特氏菌显色培养基平板		检测类别 Test Classification		委托检测		
样品标识 Sample ID		批号: 260511 保质期: 90天 生产企业: 无锡市赛微生物技术有限公司		样品编号 Sample Reference No.		CAIQS0026002463003		
包装情况 Packing Condition		包装完好		样品数量 Sample Quantity		10皿/包*1包		
接收日期 Received Date		2026年05月14日		样品性状 Sample Character		固体		
检测日期 Test Date		2026年05月15日至 2026年05月21日		其他 Comments		/		
检测方法 Method		GB 4789.28-2024						
检测项目		结果				参考范围		
基本要求	外观	平板为黄色固体, 均一性良好				-		
	pH (灭菌后)	7.20				7.2±0.2		
	功能分类	质控指标	培养条件	质控菌株	结果	特征反应	质控评定标准	特征反应
	微生物学要求	生长率		单核细胞增生李斯特氏菌 ATCC 19115	1.17	蓝绿色菌落, 有白色晕环	PR≥0.5	蓝绿色菌落, 有白色晕环
微生物学要求	选择性分离	特异性	36℃ 48h	英诺克李斯特氏菌 ATCC 33090	-	蓝绿色菌落, 无白色晕环	-	蓝绿色菌落, 无白色晕环
		选择性		大肠埃希氏菌 CICC 25012	0	-	G≤1	-
				粪肠球菌 CICC 25019	0	-		-
	..... 以下为空白 Blank below							
审核人 Assessor		张璐		批准人 Authority		薛晓晶		